

Value Assignment and Uncertainty Evaluation for Samples Used in VDSP Interlaboratory and Commutability Studies

Antonio Possolo & Blaza Toman

Statistical Engineering Division
Information Technology Laboratory

November 14, 2013



National Institute of Standards and Technology

VDSP Single Donor Samples

ANALYTICAL PROCEDURES

- For each of 50 single donor samples of serum, and for each of three analytes:
 - Ghent did one injection per sample work-up and prepared three sample work-ups
 - NIST did two injections per sample work-up, and prepared two sample work-ups

VDSP Single Donor Samples

LIMITS OF QUANTITATION


- For the concentrations of 25(OH)D₂, 25(OH)D₃, and 3-epi-25(OH)D₃:
 - Ghent: 0.6 ng mL⁻¹
 - NIST: 0.5 ng g⁻¹ for 25 samples measured by one chemist, and 1.0 ng g⁻¹ for the other 25 samples measured by another chemist

VDSP Single Donor Samples

UNITS OF MEASUREMENT

- Determinations made by NIST converted from ng g^{-1} to ng mL^{-1} using values of serum density measured by Ghent
- All values, from Ghent and from NIST, converted from ng mL^{-1} to nmol L^{-1} by multiplication by
 - 2.423 (for $25(\text{OH})\text{D}_2$)
 - 2.496 (for $25(\text{OH})\text{D}_3$ and for $3\text{-epi-}25(\text{OH})\text{D}_3$)

VDSP Single Donor Samples — Total 25(OH)D



VDSP Single Donor Samples

VALUE ASSIGNMENT

- Determined by agreement between relevant parties:
 - Values of $25(\text{OH})\text{D}_2$ and $25(\text{OH})\text{D}_3$ assigned to each sample: average of averages of determinations made by Ghent and by NIST
 - Total concentration of $25(\text{OH})\text{D}$: calculated separately for each lab as sum of lab's average concentrations of $25(\text{OH})\text{D}_2$ and $25(\text{OH})\text{D}_3$
Considering also the values for $25(\text{OH})\text{D}_2$ that were below the LoQ
 - Values assigned to $3\text{-epi-}25(\text{OH})\text{D}_3$: averages of determinations made by NIST only

VDSP Single Donor Samples

BIAS ASSESSMENT

- **Biases** estimated by differences between averages of replicated determinations of same measurand in NIST SRM 972
- None of the biases differ significantly from 0, either statistically or substantively

VDSP Single Donor Samples

REFERENCE MATERIALS USED FOR CALIBRATION

- Reference materials used for calibration, quality assurance, and bias assessment all from NIST SRM 972 — **uncertainties associated with reference values need to be propagated and expressed**

VDSP Single Donor Samples


UNCERTAINTY EVALUATION (1/2)

For each single donor sample:

- Compute averages \bar{x}_G and \bar{x}_N
- Compute associated standard uncertainties
 $u(\bar{x}_G) = s_G/\sqrt{n_G}$ and $u(\bar{x}_N) = s_N/\sqrt{n_N}$
- Assigned value $x = \frac{1}{2}(\bar{x}_G + \bar{x}_N)$
- Uncertainty component attributable to within-laboratory dispersion of values $u_{WD}(x) = \frac{1}{2}\sqrt{u^2(\bar{x}_G) + u^2(\bar{x}_N)}$
- Corresponding number of degrees of freedom given approximately by Welch-Satterthwaite formula

VDSP Single Donor Samples

UNCERTAINTY CONTRIBUTION FROM CALIBRATION







VDSP Single Donor Samples

UNCERTAINTY EVALUATION (2/2)

- Component of uncertainty $u_C(x)$ attributable to calibration obtained by interpolation from relationship between uncertainty and value for SRMs 972 and 972a
- Combined standard measurement uncertainty $u(x) = \sqrt{u_{WD}^2(x) + u_C^2(x)}$, with corresponding number of degrees of freedom ν computed using Welch-Satterthwaite formula
- Approximate 95 % coverage interval for the value of the measurand is $x \pm t_{\nu,0.975}u(x)$

VDSP Single Donor Samples

RESULTS



VDSP Commutability Study — 25(OH)D₃

